

FORM MD-10

[See sub-rule (1) rule 25]

सत्यमेव जयते

Loan Licence to Manufacture for Sale or for Distribution of Class C or Class D medical device

Loan Licence Number: MFG/MD/2020/000275

1. M/s SMB Corporation Of India, 13,33-36 Prem Industrial Estate, Subhash Road, Jogeshwari(E), Mumbai Mumbai, --, Maharashtra (India) - 400060 Telephone No.: 2228345637 FAX: 2228215157 has been licenced to manufacture for sale or for distribution the below listed medical device(s) at the premises situated at M/s SMB Medical Devices Pvt Ltd, 156 GIDC UMBERGAON, Valsad, Gujarat (India) - 396170 Telephone No.: 02602560209 FAX: 022 28215157

2. Details of medical device(s) [Annexed]

3. The names, qualifications and experience of the competent technical staff responsible for the manufacture and testing of the above mentioned medical device(s): As per records maintain by the manufacturer

4. This licence is subject to the provisions of the Medical Devices Rules, 2017 and conditions prescribed therein.

S.No.	CDSCO Details Of Device(s) SCO
1	Generic Name:Copper T 380A (TCu 380A/CuT 380A) IUD
	Model No.:Copper T 380A (TCu 380A/CuT 380A) IUD - WHO UNFPA Specification/ISO
	7439/NDA 18-680/MOH & FW New Delhi
	Intended Use:Intrauterine contraception in the woman of Child Bearing Age
	Class of medical device: Class D
	Class of medical device:Class D Material of construction:Plastic and Copper Dimension(if any):Copper T 380A : 36 mm + 5% and width 32 mm + 5%
	Dimension(if any):Copper T 380A : 36 mm \pm 5% and width 32 mm \pm 5%
	Shelflife:Up to 7 years
	Sterile or Non sterile:Sterilized
	Brand Name(if registered under the Trade Marks Act, 1999):SMB
2	Generic Name:Cu 375 IUD
	Model No.:Cu 375 IUD Sleek - WHO UNFPA Technical Specification /ISO 7439 /MOH & FW
	Specifications of Govt. Of India ,Cu 375 IUD Standard - WHO UNFPA Specification/ISO
	7439/MOH & FW New Delhi
	Intended Use:Intrauterine contraception in the woman of Child bearing age
	Class of medical device:Class D
	Material of construction:Plastic and Copper
	Dimension(if any):Standard :Vertical stem Length :36 mm \pm 5% and width 18.7 mm \pm 5%,
	Sleek :Vertical stem Length :29.5 mm \pm 5% and width 18.7 mm \pm 5%,
	Shelflife:5 Years
	Sterile or Non sterile:Sterilized
	Brand Name(if registered under the Trade Marks Act, 1999):SMB

ANNEXURE

3	Generic Name:Cu 250 IUD
	Model No.:Cu 250 IUD Standard - ISO 7439/Inhouse ,Cu 2650 IUD Sleek - ISO 7439/Inhouse
	Intended Use:Intrauterine contraception in the woman of Child Bearing age
	Class of medical device:Class D
	Material of construction:Plastic and Copper
	Dimension(if any):Standard :Vertical stem Length :36 mm \pm 5% and width 18.7 mm \pm 5%,
	Sleek :Vertical stem Length :29.5 mm ± 5% and width 18.7 mm ± 5%,
	Shelflife:5 Years
	Sterile or Non sterile:Sterilized
	Brand Name(if registered under the Trade Marks Act, 1999):SMB
4	Generic Name:TCu 380Ag IUD
	Model No.:TCu 380Ag IUD Mini - Inhouse/ISO 7439 ,TCu 380Ag IUD Normal - Inhouse/ISO
	7439 ,TCu 380Ag IUD Maxi - Inhouse/ISO 7439
	Intended Use:Intrauterine Contraception in the woman of Child bearing age
	Class of medical device:Class D
	Material of construction:Plastic and Copper-Silver alloy wire
	Dimension(if any): Mini : Length 30.5 mm \pm 5% and width 24 mm \pm 5% ,Normal - Length 33
	mm \pm 5% and width 31 mm \pm 5%, Maxi- Length 38mm \pm 5% and width 36mm \pm 5%
	Shelflife:5 Years
	Sterile or Non sterile:Sterilized
	Brand Name(if registered under the Trade Marks Act, 1999):SMB
	Generic Name:Copper T 380A (TCu 380A/CuT380A) IUD with Safe Load and Probe
5	Model No.:Copper T 380A (TCu 380A/CuT380A) IUD With safe Load and Probe - WHO UNFPA
	Specification/ISO 7439/NDA 18-680/MOH & FW New Delhi
	Intended Use:Intrauterine contraception in the woman of Child bearing age
	Class of medical device:Class D
	Material of construction:Plastic and Copper
	Dimension(if any):Copper T 380A : 36 mm \pm 5% and width 32 mm \pm 5%
	Shelflife:Up to 7 years
	Stelle of Nor stelle. Stellized
	Brand Name(if registered under the Trade Marks Act, 1999):SMB
6	Generic Name: Tubal Rings
	Model No.:Tubal Ring/Fallopian Tube Ring - IS 13009
	Intended Use:Contraception in woman
	Class of medical device:Class C
	Material of construction:Silicon Rubber
	Dimension(if any):ID. =1.0 +/_ 0.1 mm, OD = 4.0 mm ± 0.3 mm,Thickness= 2.5 mm± 0.0.15 mm
	Shelflife:4 Years
	Sterile or Non sterile:Sterilized
	Brand Name(if registered under the Trade Marks Act, 1999):SMB

Place:

Date04-Nov-20

Central Licensing Authority